



Basic information	
<b>2023/0005(COD)</b> COD - Ordinary legislative procedure (ex-codecision procedure) Regulation	Procedure completed
Transitional provisions for certain medical devices and in vitro diagnostic medical devices  Amending Regulation 2017/745 <a href="#">2012/0266(COD)</a> Amending Regulation 2017/746 <a href="#">2012/0267(COD)</a>  <b>Subject</b>  2.10.03 Standardisation, EC/EU standards and trade mark, certification, compliance 3.40.11 Precision engineering, optics, photography, medical 4.20.05 Health legislation and policy 4.60.08 Safety of products and services, product liability	

Key players				
European Parliament	<b>Committee responsible</b>		<b>Rapporteur</b>	<b>Appointed</b>
	<a href="#">ENVI</a> Environment, Climate and Food Safety			
	<b>Committee for opinion</b>		<b>Rapporteur for opinion</b>	<b>Appointed</b>
	<a href="#">EMPL</a> Employment and Social Affairs		The committee decided not to give an opinion.	
	<a href="#">IMCO</a> Internal Market and Consumer Protection		The committee decided not to give an opinion.	
Council of the European Union	<b>Council configuration</b>		<b>Meetings</b>	<b>Date</b>
	Education, Youth, Culture and Sport		6785	2023-03-07
European Economic and Social Committee				
European Committee of the Regions				

Key events			
Date	Event	Reference	Summary
06/01/2023	Legislative proposal published	COM(2023)0010 	Summary
26/01/2023	Committee referral announced in Parliament, 1st reading		
16/02/2023	Decision by Parliament, 1st reading	T9-0052/2023	Summary
16/02/2023	Results of vote in Parliament		

07/03/2023	Act adopted by Council after Parliament's 1st reading		
15/03/2023	Final act signed		
20/03/2023	Final act published in Official Journal		

Technical information	
<b>Procedure reference</b>	2023/0005(COD)
<b>Procedure type</b>	COD - Ordinary legislative procedure (ex-codecision procedure)
<b>Procedure subtype</b>	Legislation
<b>Legislative instrument</b>	Regulation
	Amending Regulation 2017/745 <a href="#">2012/0266(COD)</a> Amending Regulation 2017/746 <a href="#">2012/0267(COD)</a>
<b>Legal basis</b>	Rules of Procedure EP 170 Treaty on the Functioning of the EU TFEU 168-p4 Treaty on the Functioning of the EU TFEU 114
<b>Mandatory consultation of other institutions</b>	<a href="#">European Economic and Social Committee</a> <a href="#">European Committee of the Regions</a>
<b>Stage reached in procedure</b>	Procedure completed
<b>Committee dossier</b>	ENVI/9/11070

Documentation gateway				
<b>European Parliament</b>				
Document type	Committee	Reference	Date	Summary
Text adopted by Parliament, 1st reading/single reading		<a href="#">T9-0052/2023</a>	16/02/2023	<a href="#">Summary</a>
<b>Council of the EU</b>				
Document type	Reference	Date	Summary	
Draft final act	00001/2023/LEX	15/03/2023		
<b>European Commission</b>				
Document type	Reference	Date	Summary	
Legislative proposal	<a href="#">COM(2023)0010</a> 	06/01/2023	<a href="#">Summary</a>	
<b>Other institutions and bodies</b>				
Institution/body	Document type	Reference	Date	Summary
ESC	Economic and Social Committee: opinion, report	<a href="#">CES0203/2023</a>	24/01/2023	

## Meetings with interest representatives published in line with the Rules of Procedure

### Other Members

--	--	--

Name	Date	Interest representatives
LUENA César	14/02/2023	Edwards Lifesciences

Final act
<a href="#">Regulation 2023/0607</a> <a href="#">OJ L 080 20.03.2023, p. 0024</a>

## Transitional provisions for certain medical devices and in vitro diagnostic medical devices

2023/0005(COD) - 06/01/2023 - Legislative proposal

PURPOSE: to ensure that patients across Europe have access to safe medical devices.

PROPOSED ACT: Regulation of the European Parliament and of the Council.

ROLE OF THE EUROPEAN PARLIAMENT: the European Parliament decides in accordance with the ordinary legislative procedure and on an equal footing with the Council.

BACKGROUND: Regulations [\(EU\) 2017/745](#) (MDR) and [\(EU\) 2017/746](#) (IVDR) of the European Parliament and of the Council establish a new regulatory framework to ensure the proper functioning of the internal market for medical devices and in vitro diagnostic medical devices, based on a high level of health protection for patients and users.

In addition, the two regulations :

- set high quality and safety standards for medical devices and in vitro diagnostic medical devices in order to address the common safety issues related to these devices
- significantly strengthen key aspects of the previous regulatory framework such as the supervision of notified bodies, conformity assessment procedures, clinical evidence requirements, vigilance and market surveillance, while introducing provisions to ensure transparency and traceability of medical devices and in vitro diagnostic medical devices.

BACKGROUND: Regulations [\(EU\) 2017/745](#) and [\(EU\) 2017/746](#) of the European Parliament and of the Council establish a new regulatory framework to ensure the smooth functioning of the internal market as regards medical devices and in vitro diagnostic medical devices, taking as a base a high level of protection of health for patients and users.

In addition, the two Regulations:

- set high standards of quality and safety for medical devices and in vitro diagnostic medical devices in order to meet common safety concerns as regards such devices;
- significantly reinforce key elements of the previous regulatory framework, such as the supervision of notified bodies, conformity assessment procedures, clinical evidence requirements, vigilance and market surveillance, whilst introducing provisions ensuring transparency and traceability regarding medical devices and in vitro diagnostic medical devices.

Due to the impact of the COVID-19 pandemic, the date of application of the MDR was postponed by one year to 26 May 2021 by [Regulation \(EU\) 2020/561](#) of the European Parliament and of the Council, while the date of 26 May 2024 was maintained as end of the transition period by which certain devices that continue to comply with Directive 90/385/EEC or Directive 93/42/EEC may be placed on the market or put into service.

Despite considerable progress over the past years, the overall capacity of conformity assessment ('notified') bodies remains insufficient to carry out the tasks required of them. In addition, many manufacturers are not sufficiently prepared to meet the strengthened requirements of the MDR by the end of the transition period. This is threatening the availability of medical devices on the EU market.

The overall goal of the proposed amendments is to maintain patients' access to a wide range of medical devices while ensuring the transition to the new framework.

CONTENT: this proposal does not alter the MDR or IVDR in substance, nor does it impose any new obligations on the parties concerned. The main purpose of this proposal is to amend the transitional provisions, allowing for an **additional period of time to transition to the MDR's requirements** to avoid shortages.

In concrete terms, the proposal therefore aims to:

- **extend the current transition period** in Article 120 of the MDR, subject to certain conditions, so that only those devices that are safe and for which manufacturers have already taken steps to transition to the MDR will benefit from the additional time. The transition period would be extended from 26 May 2024 to **31 December 2027** for higher risk devices and to **31 December 2028** for lower and medium risk devices;
- **delete the 'sell off'** deadline in the relevant provisions of the MDR and IVDR, i.e. the date until which devices that are placed on the market before or during the transition period and are still in the supply chain when the extended transition period is over can be made available.

The extension of the transition period is complemented by an **extension of the validity of certificates** issued under the previous Council Directives 90/385/EEC and 93/42/EEC for the devices benefiting from the extended transition period. Also the validity of certificates that have already expired since 26 May 2021 would be extended, subject to certain conditions.

## Transitional provisions for certain medical devices and in vitro diagnostic medical devices

2023/0005(COD) - 16/02/2023 - Text adopted by Parliament, 1st reading/single reading

The European Parliament adopted by 537 votes to 3, with 24 abstentions, a legislative resolution on the proposal for a regulation of the European Parliament and of the Council amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices.

Parliament adopted its position at first reading under the ordinary legislative procedure by taking over the Commission proposal.

The aim of the regulation is to **address the risks of shortages of medical devices and in vitro diagnostic medical devices** and thus maintain patient access to a wide range of medical devices.

Regulations (EU) 2017/745 (MDR) and (EU) 2017/746 (IVDR) of the European Parliament and of the Council establish a new regulatory framework to ensure the proper functioning of the internal market for medical devices and in vitro diagnostic medical devices, based on a high level of health protection for patients and users.

The MDR is applicable from 26 May 2021. The transition period provided for in the Regulation will end on 26 May 2024.

The IVDR is applicable from 26 May 2022. A staggered extension of its transition period, from 26 May 2025 for high-risk in vitro diagnostics to 26 May 2027 for low-risk in vitro diagnostics and to 26 May 2028 for certain provisions concerning devices manufactured and used in health care facilities, has been adopted.

Despite the steady increase in the number of notified bodies designated in accordance with Regulation (EU) 2017/745, the overall capacity of notified bodies is still not sufficient to ensure the conformity assessment of the large number of devices covered by certificates issued in accordance with Directive 90/385/EEC or Directive 93/42/EEC before 26 May 2024. A large number of manufacturers, especially small and medium-sized enterprises, are not sufficiently prepared to demonstrate compliance with the requirements of Regulation (EU) 2017/745.

The amending Regulation therefore:

- **extends the current transition period** in Article 120 of the MDR, subject to certain conditions, so that only those devices that are safe and for which manufacturers have already taken steps to transition to the MDR will benefit from the additional time. The transition period would be extended from 26 May 2024 to 31 December 2027 for higher risk devices and to **31 December 2028** for lower and medium risk devices;
- **deletes the 'sell off' deadline** in the relevant provisions of the MDR and IVDR, i.e. the date until which devices that are placed on the market before or during the transition period and are still in the supply chain when the extended transition period is over can be made available.
- **extends the validity of certificates** issued under the previous Council Directives 90/385/EEC and 93/42/EEC for the devices benefiting from the extended transition period. Also, the validity of certificates that have already expired since 26 May 2021 would be extended, subject to certain conditions.